

FLEXCELL® FLEXSTOP™

(cat. no. BFS-3000)

Be Selective! Use a FlexStop to block stretching

- Reusable valved rubber stopper inserts into underside of BioFlex, Tissue Train and UniFlex culture plate wells to prevent pressure-induced deformation.
- Can be used as a negative control (static culture) when testing mechanical load effects on cells in BioFlex, Tissue Train and UniFlex culture systems.



FlexStop includes:

- 12 Rubber stoppers
- 12 Brass pins for insertion into the BioFlex, TissueTrain and UniFlex wells




Figure 11. Schematic of FlexStop mechanism to prevent vacuum-induced deformation of well bottom.